

**DATE RECEIVED:** FEBRUARY 18, 2021 **DATE REPORTED:** MARCH 12, 2021

**PAGE:** 1 of 5

Test Report No. : CPSA/210358305-CA58514

Company : Maxter Glove Manufacturing Sdn Bhd

Lot 6070, Jalan Haji Abdul Manan

6<sup>th</sup> Miles Off Jalan Meru

41050 Klang, Selangor, Malaysia

## TEST REPORT

Sample Description : Non Sterile Powder Free Latex Examination Gloves

Lot No. : 61062 03652

Manufacturer : Maxter Glove Manufacturing Sdn Bhd

Size : Medium Quantity Tested : 200 pieces

Test Conducted : Freedom from holes Test Method : EN455 Part 1:2000

Testing Period : 18 February 2021 - 08 March 2021

Based on submitted samples, the following results obtained :-

Acceptable Quality Limit (AQL): 1.5 Accept: 7 Found: 0

Result : Within AQL

Remark: This remark supersedes report no. CPSA/210257057-CA58514

SIGNED FOR AND ON BEHALF OF SGS (MALAYSIA) SDN BHD

CHEE TUCK CHOON SECTION HEAD

IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="https://www.sgs.com/en/terms-and-conditions">https://www.sgs.com/en/terms-and-conditions</a> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="https://www.sgs.com/en/terms-and-conditions/terms-e-document.">https://www.sgs.com/en/terms-and-conditions</a> at the time of its intervention of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.



**DATE RECEIVED:** FEBRUARY 18, 2021 **DATE REPORTED:** MARCH 12, 2021

**PAGE:** 2 of 5

Test Report No. : CPSA/210358305-CA58514

Company : Maxter Glove Manufacturing Sdn Bhd

Lot 6070, Jalan Haji Abdul Manan

6<sup>th</sup> Miles Off Jalan Meru

41050 Klang, Selangor, Malaysia

## TEST REPORT

Sample Description : Non Sterile Powder Free Latex Examination Gloves

Lot No. : 61062 03652

Manufacturer : Maxter Glove Manufacturing Sdn Bhd

Size : Medium
Quantity Tested : 13 pieces
Test Conducted : Dimensions

Test Method : EN 455 Part 2:2015

Testing Period : 18 February 2021 - 08 March 2021

Based on submitted samples, the following results obtained :-

Size	M	M	M	M	M	M	M	M	M	M	M	M	M	Median
Width Median: 95±10mm	94	93	95	93	94	94	93	94	94	93	95	94	93	94
Length Median: ≥ 240mm	250	249	253	252	248	247	250	251	248	248	252	252	245	250

Remark: This remark supersedes report no. CPSA/210257057-CA58514

SIGNED FOR AND ON BEHALF OF SGS (MALAYSIA) SDN BHD

CHEE TUCK CHOON SECTION HEAD

IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="https://www.sgs.com/en/terms-and-conditions">https://www.sgs.com/en/terms-and-conditions</a> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="https://www.sgs.com/en/terms-and-conditions/terms-e-document.">https://www.sgs.com/en/terms-and-conditions</a> at the time of its intervention of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.



**DATE RECEIVED:** FEBRUARY 18, 2021 **DATE REPORTED:** MARCH 12, 2021

**PAGE:** 3 of 5

Test Report No. : CPSA/210358305-CA58514

**Company** : Maxter Glove Manufacturing Sdn Bhd

Lot 6070, Jalan Haji Abdul Manan

6th Miles Off Jalan Meru

41050 Klang, Selangor, Malaysia

## TEST REPORT

Sample Description Non Sterile Powder Free Latex Examination Gloves

Lot No. 61062 03652

Manufacturer Maxter Glove Manufacturing Sdn Bhd

Size Medium

Sample Quantity 13 pieces per test

Test Conducted Force at Break During Shelf Life and After Challenge

Test Method EN 455 Part 2:2015 Ageing  $70 \pm 2$  Deg C for 168 hrs

**Testing Period** 18 February 2021 - 08 March 2021

#### Force at Break, N

SIZE	SAMPLE NO.	BEFORE AGING	AFTER AGING
M	1	9.3	7.5
	2	10.2	7.0
	3	8.5	8.2
	4	9.1	7.0
	5	8.8	7.4
	6	9.8	8.1
	7	9.3	7.8
	8	9.4	7.5
	9	8.5	8.0
	10	8.5	7.8
	11	9.3	8.5
	12	9.0	7.5
	13	9.3	8.0
Median		9.3	7.8
Requirement		≥ 6.0	≥ 6.0

Remark: This remark supersedes report no. CPSA/210257057-CA58514

SIGNED FOR AND ON BEHALF OF SGS (MALAYSIA) SDN BHD

CHEE TUCK CHOON SECTION HEAD

IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at https://www.sgs.com/en/terms-and-conditions and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at https://www.sgs.com/en/terms-and-conditions/terms-e-document. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.



**DATE RECEIVED:** FEBRUARY 18, 2021 **DATE REPORTED:** MARCH 12, 2021

**PAGE:** 4 of 5

Test Report No. : CPSA/210358305-CA58514

Company : Maxter Glove Manufacturing Sdn Bhd

Lot 6070, Jalan Haji Abdul Manan

6<sup>th</sup> Miles Off Jalan Meru

41050 Klang, Selangor, Malaysia

## TEST REPORT

Sample Description : Non Sterile Powder Free Latex Examination Gloves

Lot No. : 61062 03652

Manufacturer : Maxter Glove Manufacturing Sdn Bhd

Size : Medium
Quantity Tested : 5 pieces
Test Conducted : Powder Content
Test Method : EN455 Part 3:2015

Testing Period : 18 February 2021 - 08 March 2021

On testing the samples, the following results were obtained: -

SIZE Average Powder Mass per Glove

M 1.26 mg

Remark: This remark supersedes report no. CPSA/210257057-CA58514

SIGNED FOR AND ON BEHALF OF SGS (MALAYSIA) SDN BHD

CHEE TUCK CHOON SECTION HEAD

IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="https://www.sgs.com/en/terms-and-conditions">https://www.sgs.com/en/terms-and-conditions</a> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="https://www.sgs.com/en/terms-and-conditions/terms-e-document.">https://www.sgs.com/en/terms-and-conditions</a> at the time of its intervention of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.



**DATE RECEIVED:** FEBRUARY 18, 2021 **DATE REPORTED:** MARCH 12, 2021

**PAGE:** 5 of 5

Test Report No. : CPSA/210358305-CA58514

Company : Maxter Glove Manufacturing Sdn Bhd

Lot 6070, Jalan Haji Abdul Manan

6<sup>th</sup> Miles Off Jalan Meru

41050 Klang, Selangor, Malaysia

# **TEST REPORT**

Sample Description : Non Sterile Powder Free Latex Examination Gloves

Lot No. : 61062 03652

Manufacturer : Maxter Glove Manufacturing Sdn Bhd

Size : Medium Quantity Tested : 8 pieces

Test Conducted : Aqueous Extractable Protein Content

Test Method : EN455 Part 3:2015

Testing Period : 18 February 2021 - 08 March 2021

## EXTRACTABLE PROTEIN CONTENT

Analysis had been carried out as per your request. We report the following results:

SIZE	AQUEOUS EXTRACTABLE PROTEIN CONTENT, μg/g					
M	14					

Note: Detection limit =  $10 \mu g/g$ 

Remark: This remark supersedes report no. CPSA/210257057-CA58514

SIGNED FOR AND ON BEHALF OF SGS (MALAYSIA) SDN BHD

\*\*\* End of Report \*\*\*

CHEE TUCK CHOON SECTION HEAD

IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="https://www.sgs.com/en/terms-and-conditions">https://www.sgs.com/en/terms-and-conditions</a> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="https://www.sgs.com/en/terms-and-conditions/terms-e-document.">https://www.sgs.com/en/terms-and-conditions/terms-e-document.</a> Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.